



**Health &
Safety
Executive**



BASEEFA

British Approvals Service for Electrical Equipment in Flammable Atmospheres

CERTIFICATE OF CONFORMITY VARIATION

THIS IS TO CERTIFY THAT CERTIFICATE OF CONFORMITY BAS NO Ex 87B2087X

Issued to **ROSEMOUNT ANALYTICAL, UNILOC DIVISION
of Irvine, California, USA**


for the **MODEL 1181T CONDUCTIVITY TRANSMITTER**

is hereby extended to apply to apparatus designed and constructed in accordance with the specification set out in the Schedule of the said Certificate but having the variations specified in the following Schedule.

VARIATION ONE To permit minor modifications to the housing (casting) drawing, which do not affect Intrinsic Safety.

| <u>DRAWING Number</u> | <u>Issue</u> | <u>Date</u> | <u>Description</u> |
|---------------------------|--------------|-------------|--------------------|
| 32495 sheets 1 & 2 | B | 29.6.87 | Housing (casting) |

Code for Ex 87B2087X/1 : EEx ia IIB T4 File: EECS 0911/02/005
(Tamb = 55°C)
(as original)



CERTIFICATE OF CONFORMITY BAS NO Ex 87B2087X/1

95/03

B HILL
DIRECTOR
Dated 25 November 1987