



British Approvals Service for Electrical
Equipment in Flammable Atmospheres

Certificate of Conformity Variation

THIS IS TO CERTIFY THAT CERTIFICATE BAS NO Ex 87B2087X

Held by ROSEMOUNT ANALYTICAL INC
of Irvine, California, USA

for the MODEL 1181T CONDUCTIVITY TRANSMITTER

is hereby extended to apply to apparatus designed and constructed in accordance with the specification set out in the Schedule of the said Certificate but having the variations specified in the following Schedule.

A copy of this Supplementary Certificate shall be attached to the original Certificate.

VARIATION TEN To permit capacitor C33 to change in value from 0.01 μ F to 0.033 μ F.

DRAWINGS

Number	Issue	Date	Description
2400204	H	4.29.94	Schematic
22934-01	G	4.19.94	Signal Input Board

File No: EECS 0911/02/005

Addendum No 1 dated 16 May 1994
to Test Report No 87(i)192



CERTIFICATE BAS NO Ex 87B2087X/10

LP

I M CLEARE
DIRECTOR EECS
23 May 1994

This certificate is issued under NACCB accreditation No. 020



Electrical Equipment Certification Service
Health and Safety Executive
Harpur Hill, Buxton, Derbyshire, SK17 9JN, United Kingdom
Tel: 0298 26211 Fax: 0298 79514 Telex: 668113 RLSD G

