

British Approvals Service for Electrical  
Equipment in Flammable Atmospheres



## *Certificate of Conformity - Variation*

SUPPLEMENTARY CERTIFICATE BAS No. Ex 96D2494/4

This is to certify that Apparatus Certificate number:

**Ex 96D2494**

held by:

**ROSEMOUNT ANALYTICAL INC**

of:

**2400 Barranca Parkway, Irvine, California 92714-5018, USA**

for the:

**pH/ORP TRANSMITTER MODEL 3081**

is hereby extended to apply to the apparatus designed and constructed in accordance with the specification set out in the Schedule of the said Certificate but having any variations specified in the Schedule attached to this certificate and the documents therein referred to.

This Supplementary Certificate shall be held with the original Certificate.

File No: EECS 0911/02/015

Sheet 1 of 2

This certificate is granted subject to the general conditions of the Electrical Equipment Certification Service. It does not necessarily indicate that the apparatus may be used in particular industries or circumstances. Representation of equipment as "Certified" is valid only when the number of the prime certificate to which this certificate is a supplement is given on the relevant EECS Manufacturing Licence or Verification Certificate.



**I M CLEARE**  
**DIRECTOR**  
27 July 1998



Registration Number  
020  
The use of the Accreditation  
Mark indicates accreditation in  
respect of those activities  
covered by the accreditation  
certificate number 020.

  
**HSE**  
Health & Safety  
Executive

**Electrical Equipment Certification Service**  
Health and Safety Executive  
Harpur Hill, Buxton, Derbyshire, SK17 9JN, United Kingdom  
Tel: 01298 28000 Fax: 01298 28244



**Supplementary Certificate BAS No. Ex 96D2494/4**

**VARIATION FOUR**

To permit the following:-

1. Changes to the CPU printed circuit board which include:
  - a) Minor correction to pin designation on connector JP4
  - b) Minor changes to production details
  - c) Alternative pin configuration on various jumper connectors.
2. The use of an alternative software program in the previously certified models thus creating a new Model 3081FG.

The above changes do not affect intrinsic safety.

**DRAWINGS**

<u>Number</u>	<u>Issue</u>	<u>Date</u>	<u>Description</u>
23641-02/04	C	6.5.98	PCB assembly, CPU
9241337-00	A	Jul 13 98	Label - Model 3081 FG