

EC DECLARATION OF CONFORMITY

Form 6

We, Emerson Process Management GmbH & Co. OHG, Industriestraße 1, D-63594 Hasselroth, Germany

Herewith declare that the product,

X-STREAM Oxygen Remote Interface, Model Xi

manufactured by, Rosemount Analytical Inc., 6565P Davis Industrial Parkway, Solon, Ohio, 44139, USA

Is in conformity with the provisions of:

- The EMC Directive (2004/108/EC) including the latest amendments with the application of the harmonized standards:
EN 61326:2006 *Electrical equipment for measurement, control and laboratory use, EMC requirements, Part 1: General requirements*
- The Low Voltage Directive (2006/95/EC) including the latest amendments with the application of the harmonized standards:
EN 61010-1:2001 *Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use. Part 1: General Requirements*
- The Pressure Equipment Directive (97/23/EC) including the latest amendments with the application of the harmonized standards:
This equipment has been designed and manufactured with sound engineering practices in accordance with Article 3, Paragraph 3 of the PED

Solon, OH USA
(Place of Issue)

21 July 2009
(Date of Issue)

RAIO EC-33/3
(Document No.)



Andy Kemish, VP Analytical Europe, Analytical Group