

EC DECLARATION OF CONFORMITY

Document number: RAE/X-STREAM-E10

Date: June 2009

We,

Emerson Process Management GmbH & Co. OHG

located at

Industriestrasse 1, D-63594 Hasselroth, Germany

declare under our sole responsibility that our gas analyzer, type

X-STREAM

to which this declaration relates is in conformity with the provisions of:

2004/108/EC EMC Directive

with the application of the harmonized standards including the latest amendments:

EN 61326-1:2006 Electrical equipment for measurement, control and laboratory use – EMC requirements

2006/95/EC Low Voltage Directive

with the application of the harmonized standards:

EN 61010-1:2001 Safety requirements for electrical equipment for measurement, control and laboratory use Part 1: General requirements

97/23/EC Pressure Equipment Directive

This analyzer has been designed and manufactured considering article 3, paragraph 3 of the above mentioned directive and therefore CE marking does not refer to this directive.

This document covers all 19“ and fieldhousing X-STREAM gas analyzer variations.

Hasselroth, June 2009



.....
(Signature)

Andy Kemish

(Name)

VP Rosemount Analytical Europe

(Function name)

ROSEMOUNT
Analytical



EMERSON
Process Management

This declaration confirms the compliance with announced directives but does not include the assurance of properties.
The safety and installation instructions of the documentation have to be followed.