

# EC DECLARATION OF CONFORMITY

Document number: RAE/X-STREAM 2 DC-E1

Date: March 2008

We,

**Emerson Process Management GmbH & Co. OHG**

located at

Industriestrasse 1, D-63594 Hasselroth, Germany

declare under our sole responsibility that our gas analyzer, type

**X-STREAM 2**

to which this declaration relates is in conformity with the provisions of:

---

## 2004/108/EG EMC Directive

with the application of the harmonized standards:

EN 61326-1:2006      Electrical equipment for measurement, control and laboratory use -  
EMC requirements

---


## 97/23/EC Pressure Equipment Directive

This analyzer has been designed and manufactured considering article 3, paragraph 3 of the above mentioned directive and therefore CE marking does not refer to this directive.

---

This document covers all ½ 19“ X-STREAM 2 gas analyzer variations with DC power supply.

Hasselroth, March 2008

  
.....  
(Signature)

Andy Kemish

(Name)

VP Rosemount Analytical Europe

(Function name)

**ROSEMOUNT**<sup>®</sup>  
Analytical

This declaration confirms the compliance with announced directives but does not include the assurance of properties.  
The safety and installation instructions of the documentation have to be followed.

  
**EMERSON**  
Process Management