

EC DECLARATION OF CONFORMITY

Document number: RAE/X-STREAM X2 AC-E1

Date: July 2009

We,

Emerson Process Management GmbH & Co. OHG

located at

Industriestrasse 1, D-63594 Hasselroth, Germany

declare under our sole responsibility that our gas analyzer, type

X-STREAM X2

to which this declaration relates is in conformity with the provisions of:

2004/108/EC EMC Directive

with the application of the harmonized standards including the latest amendments:

EN 61326-1:2006 Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements

2006/95/EC Low Voltage Directive

with the application of the harmonized standards:

EN 61010-1:2001 Safety requirements for electrical equipment for measurement, control and laboratory use
Part 1: General requirements

Last two digits of the year in which the CE marking was affixed: 08

97/23/EC Pressure Equipment Directive

This analyzer has been designed and manufactured considering article 3, paragraph 3 of the above mentioned directive and therefore CE marking does not refer to this directive.

This document covers all general purpose 19" and fieldhousing X-STREAM X2 gas analyzer variations with AC power supply.

Hasselroth, July 2009



.....
(Signature)

Andy Kemish
(Name)

VP Rosemount Analytical Europe
(Function name)

ROSEMOUNT
Analytical



EMERSON
Process Management

This declaration confirms the compliance with announced directives but does not include the assurance of properties.
The safety and installation instructions of the documentation have to be followed.