

We,

**Emerson Process Management
GmbH & Co. OHG**

located at

Industriestrasse 1, D-63594 Hasselroth, Germany

declare under our sole responsibility that the equipment described below complies with the basic safety and health requirements of the EC directive 94/9/EC concerning design, construction and placing on the market.

Equipment covered by this declaration


Description: Gasanalyzer with non-incendive design for installation in Ex-Zone 2 areas
Type : X-STREAM FN

Protection Concept: Water and dust protected fieldhousing (IP66), internal components selected to be non-incendive, non-sparking and safe by design acc.


- DIN EN 60079-0:2004
- DIN EN 60079-15:2005
- DIN EN 13463-1:2002
- DIN EN 13463-5:2004

A representative sample of the analyzer has been approved by an independent testhouse and found to comply with the requirements. The test results are recorded in test report 05KFS106.

Marking of the equipment:



Hasselroth, February 2008



(Signature)

Andy Kemish
(Name)

VP Rosemount Analytical Europe
(Function name)

This declaration becomes void if the equipment is modified without our written permission.