



Certificate of Compliance

Certificate Number: LR 105173-1

Revision: LR 105173-7

Date Issued: May 25, 1998

Issued to: Fisher-Rosemount GmbH and Co.
Industriestrasse 1
Hasselroth, D-63594
GERMANY
Attention: Mr Uwe Schmidt

The products listed below are eligible to bear the CSA Mark
¹NOTE: The "NRTL/C" indicator also appears adjacent to the CSA Mark.

Issued by: Neil Forsyth, C.E.T.
Edmonton, AB Canada

Signature: _____

PRODUCTS

Gas Analyzers, Models NGA2000P and NGA2000PE, cord connected, rated 120/240 V, 50/60 Hz, 2.5/1.5 A.

APPLICABLE STANDARDS

- | | | |
|-------------------|-----------|---|
| CAN/CSA-C22.2 No. | 1010.1-92 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements |
| ANSI/UL Std No. | 3101.1 | - Safety Requirements for Electrical Equipment for Laboratory Use |

¹The "NRTL/C" indicator adjacent to the CSA Mark signifies that the product has been evaluated to the applicable ANSI/UL and CSA Standards, for use in the U.S. and Canada. NRTL, i.e. Nationally Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.