



CSA INTERNATIONAL

Certificate of Compliance

Certificate: 1921112

Master Contract: 185562

Project: 2149429

Date Issued: 2009/03/25

Issued to: Emerson Process Management
GmbH & Co. OHG
Industriestrasse 1
Hasselroth, 63594
Germany

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: Daniel Duguay

Authorized by: Helene Vaillancourt,
Operations Manager

PRODUCTS

- CLASS 8721 85 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards
CLASS 8721 05 - LABORATORY EQUIPMENT - Electrical

Gas analyzer module, Model: NGA 2000 CLD Analyzer Module, Rated: 24Vdc, 5A, 120W.

APPLICABLE REQUIREMENTS



Certificate: 1921112

Master Contract: 185562

Project: 2149429

Date Issued: 2009/03/25

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std No. 61010-1, Second Edition - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

CONDITIONS OF ACCEPTABILITY

The unit was evaluated as a component where the suitability of the combination must be determined with the end use product.