

2202 2nd Avenue
Regina, SK S4R 1K3
Canada

306-787-1443
info@tsask.ca
www.tsask.ca

REGISTRATION APPROVAL

Friday, July 12, 2013

ABSA
9410 - 20th Avenue NW
Edmonton, AB
T6N 0A4

ATTENTION : Cynthia Formaniuk

Our File 9071 [0 F


With reference to your submission respecting the registration of the item below, for legal use in the province, please note we have surveyed, approved and registered this design as noted.

MANUFACTURER : Rosemount Analytical Inc.	
CATALOG OR DRAWING : PDS-71-398 Rev E	
ITEM : TUPH SENSORS: MODELS 398 & 398R	CRN : OF7880.23

We wish to point out that every fitting must be constructed strictly in accordance with the registered design.

Fitting registrations are required to be resubmitted for validation after ten (10) years from the registration date in accordance with CSA B51, Clause 4.2.1. The date of expiry for this registration is **August 30, 2016**

Sincerely,



Adenike Adeoti, M.Sc. P.Eng.
Codes and Standards Compliance

REMARKS: CRN registered under reciprocal agreement and conditional upon compliance with the notes on ABSA's registration.



the pressure equipment safety authority

**STATUTORY DECLARATION
Registration of Fittings**

In this space, show facsimile of manufacturer's logo or trademark as it will appear on the fitting.

ROSEMOUNT[®]
Analytical

I, Dana Crowley,

Manager of Compliance and Test Engineering

(company title, e.g. vice president, plant manager, chief engineer) (must be in a position of authority)

of Emerson Process Management, Rosemount Analytical

(name of manufacturer)

located at 2400 Barranca Pkwy, Irvine, CA and Circuito del Progreso #27, Parque Industrial Progreso, Mexicali CP-21190

(plant address)

do solemnly declare that the fittings listed hereunder, which are subject to the Safety Codes Act (check one)

comply with the requirements of _____ which specifies the dimensions,
(title of recognized North American Standard)

materials of construction, pressure/temperature ratings and identification marking of the fittings, or

are not covered by the provisions of a recognized North American standard and are therefore manufactured to comply with Product Data Sheets as supported by the attached data which identifies the dimensions, materials of construction, pressure/temperature ratings and the basis for such ratings, and the marking of the fittings for identification.

I further declare that the manufacture of these fittings is controlled by a quality control program which has been verified by the following authority, Det Norske Veritas (DNV) as being suitable for the manufacture of these fittings to the stated standard. The fittings covered by this declaration, for which I seek registration, are category F type fittings


In support of this application, the following information, calculations and/or test data are attached:

Proof pressure test report and Product Data Sheets

DECLARED before me at Irvine in the County of Orange

this 12TH day of March, 2013
(Month) (Year)

(print) CATHERINE ANN STIP
(sign) Catherine Ann Stip
(A Commissioner for Oaths)



Technical Safety Authority
of Saskatchewan
Boiler & P.V. Safety Unit

For Office Use Only

To the best of my knowledge and belief, the application meets the requirements of the Safety Codes Act and C.S.A. Standard B51, Clause 4.2, and is accepted for registration in Category _____

Registration Number: OF 7880.23

Aug 30/2016

Date Registered: Jul 12/2013

CRN... OF 7880.23
9071
REGISTERED
Date
(For the Administrator/Chief Inspector of Alberta)
Expiry Date: Design Survey Office