


UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES


NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: Rosemount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92606	
PLANT LOCATIONS: Address above and Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY	
<p><input checked="" type="radio"/> A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers</p> <p><input type="radio"/> B Flanges: all flanges</p> <p><input type="radio"/> C Valves: all line valves</p> <p><input type="radio"/> D Expansion joints, flexible connections, and hose assemblies: all types</p> <p><input type="radio"/> E Strainers, filters, separators, and steam traps</p> <p><input type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters</p> <p><input type="radio"/> G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs</p> <p><input type="radio"/> H Pressure retaining components that do not fall into one of the above categories</p> <p><input type="radio"/> N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/>, (Meeting CNSC or ASME requirements)</p>	<p>TITLE OF THE STANDARD OF CONSTRUCTION</p> <p>ASME B31.3 and proprietary standards, supported by proof pressure tests and calculations.</p>
SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT	
	<p>TYPE OF CONSTRUCTION</p> <p>FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/></p> <p>CAST <input type="checkbox"/> OTHER <input checked="" type="checkbox"/></p> <p>DESCRIBE OTHER: Cast and machined</p>
LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED:	
<p>Drawings of main pressure retaining parts of kits 23166-00/01, 23311-00/01, 23765-00/01, and 23240-00</p> <p>Report containing thread strength calculations for the main pressure retaining parts of the kits listed above</p> <p>Proof pressure reports RAI 01 rev. 2 and RAI 02 rev.2.</p>	

DECLARATION:

I, Dana Crowley (see note 3) employed by Rosemount Analytical, Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: *Dana Crowley*
 Declared before me at Irvine, County of Orange, California
 This 30th day of November AD 2012
 Commissioner of Oaths or Notary Public: (sign) *Catherine Ann Stip*



<p>This space for Regulatory Authority use</p> <p>This registration must be revalidated after ten (10) years from the date of acceptance.</p> <p>CRN: <u>0A12814.2 Add.</u></p> <p>FID#: <u>14714</u></p> <p>Notes: 1. All fittings shall be registered in the name of the Manufacturer. 2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation. 3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product. 4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.</p>	<p>ACCEPTED</p> <p>PROVINCE OF PRINCE EDWARD ISLAND ENVIRONMENT, LABOUR & JUSTICE</p> <p>C.R.N. <u>0A12814.29 Add</u> DATE: <u>Feb 1/13</u></p> <p>INSPECTION SERVICES SECTION BOILER/PRESSURE VESSEL BRANCH</p> <p>Sect. 1.0 - Fittings Rev. 1 06/2003</p>
--	---

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK NOVA SCOTIA PRINCE EDWARD ISLAND NEWFOUNDLAND AND LABRADOR
 NUNAVUT YUKON NORTHWEST TERRITORIES

MANUFACTURERS NAME: Rosemount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92606	
PLANT LOCATIONS: Address above and Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY	
<input checked="" type="radio"/> Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers <input type="radio"/> Flanges: all flanges <input type="radio"/> Valves: all line valves <input type="radio"/> Expansion joints, flexible connections, and hose assemblies: all types <input type="radio"/> Strainers, filters, separators, and steam traps <input type="radio"/> Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters <input type="radio"/> Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs <input type="radio"/> Pressure retaining components that do not fall into one of the above categories <input type="radio"/> Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)	TITLE OF THE STANDARD OF CONSTRUCTION ASME B31.3 and proprietary standards, supported by proof pressure tests and calculations.
SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT <div style="text-align: center; font-size: 24pt; font-weight: bold;">ROSEMOUNT</div> Analytical	TYPE OF CONSTRUCTION FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/> CAST <input type="checkbox"/> OTHER <input checked="" type="checkbox"/> DESCRIBE OTHER: Cast and machined
LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED: Drawings of main pressure retaining parts of kits 23186-00/01, 23311-00/01, 23765-00/01, and 23240-00 Report containing thread strength calculations for the main pressure retaining parts of the kits listed above Proof pressure reports RAI 01 rev. 2 and RAI 02 rev.2.	

DECLARATION:


I Dana Crowley (PNS Note 3) employed by Rosemount Analytical, Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: [Signature]

Declared before me at Irvine, County of Orange, California

This 30th day of November AD 2012

Commissioner of Oaths or Notary Public: (sign) [Signature]



This space for Regulatory Authority Use

This registration must be revalidated after ten (10) years from the date of acceptance.

CRN: 0A12814.2 Add.

FID#: 14714

Date: Feb. 1/13

CRN: 0A12814-28ADD.....

Dwg: as described

Signed: Yvonne Perry

... of ...
Part

Notes:

- All fittings shall be registered in the name of the Manufacturer.
- Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
- The declaration shall be made by the person having full authority and responsibility for the quality of the end product.
- Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.

1.0 - Fittings Rev. 1 06/2003

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES


NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: Rosemount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92606	
PLANT LOCATIONS: Address above and Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY	
<p><input checked="" type="radio"/> A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers</p> <p><input type="radio"/> B Flanges: all flanges</p> <p><input type="radio"/> C Valves: all line valves</p> <p><input type="radio"/> D Expansion joints, flexible connections, and hose assemblies: all types</p> <p><input type="radio"/> E Strainers, filters, separators, and steam traps</p> <p><input type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters</p> <p><input type="radio"/> G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs</p> <p><input type="radio"/> H Pressure retaining components that do not fall into one of the above categories</p> <p><input type="radio"/> N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/>, (Meeting CNSC or ASME requirements)</p>	TITLE OF THE STANDARD OF CONSTRUCTION ASME B31.3 and proprietary standards, supported by proof pressure tests and calculations.
SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT	TYPE OF CONSTRUCTION
ROSEMOUNT[®] Analytical	FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/> CAST <input type="checkbox"/> OTHER <input checked="" type="checkbox"/> DESCRIBE OTHER: Cast and machined
LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED: Drawings of main pressure retaining parts of kits 23166-00/01, 23311-00/01, 23765-00/01, and 23240-00 Report containing thread strength calculations for the main pressure retaining parts of the kits listed above Proof pressure reports RAI 01 rev. 2 and RAI 02 rev.2.	

DECLARATION:

I, Dana Crowlev (see note 3) employed by Rosemount Analytical, Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: *Dana Crowlev*
 Declared before me at Irvine, County of Orange, California
 This 30TH day of November AD 2012
 Commissioner of Oaths or Notary Public: (sign) *Catherine Ann Stip*



This space for Regulatory Authority use
 This registration must be revalidated after ten (10) years from the date of acceptance.

CRN: 0A12814.2 Add.
 FID#: 14714

REGISTRATION ONLY
 CRN: 0A12814.2 7 Add.

EXAMINER: *Dana Crowlev*
 for
CHIEF BOILER INSPECTOR

DATE: Feb 1 / 2013

Sect. 1.0 - Fittings Rev. 1 06/2003

DLRs PVs

FITTINGS COMPONENTS

- Notes:
1. All fittings shall be registered in the name of the Manufacturer.
 2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
 3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product.
 4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES

NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: Rosemount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92606	
PLANT LOCATIONS: Address above and Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
<p>CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY</p> <p><input checked="" type="radio"/> A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers</p> <p><input type="radio"/> B Flanges: all flanges</p> <p><input type="radio"/> C Valves: all line valves</p> <p><input type="radio"/> D Expansion joints, flexible connections, and hose assemblies: all types</p> <p><input type="radio"/> E Strainers, filters, separators, and steam traps</p> <p><input type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters</p> <p><input type="radio"/> G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs</p> <p><input type="radio"/> H Pressure retaining components that do not fall into one of the above categories</p> <p><input type="radio"/> N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)</p>	<p>TITLE OF THE STANDARD OF CONSTRUCTION</p> <p>ASME B31.3 and proprietary standards, supported by proof pressure tests and calculations.</p>
<p>SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT</p> <p align="center">ROSEMOUNT Analytical</p>	<p>TYPE OF CONSTRUCTION</p> <p>FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/></p> <p>CAST <input type="checkbox"/> OTHER <input checked="" type="checkbox"/></p> <p>DESCRIBE OTHER: Cast and machined</p>
<p>LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED:</p> <p>Drawings of main pressure retaining parts of kits 23166-00/01, 23311-00/01, 23765-00/01, and 23240-00 Report containing thread strength calculations for the main pressure retaining parts of the kits listed above Proof pressure reports RAI 01 rev. 2 and RAI 02 rev. 2.</p>	

DECLARATION:


I Dana Crowley (see note 3) employed by Rosemount Analytical, Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: [Signature]

Declared before me at Irvine, County of Orange, California

This 30th day of November AD 2012

Commissioner of Oaths or Notary Public: (sign) [Signature]



This space for Regulatory Authority use


This registration must be revalidated after ten (10) years from the date of acceptance

CRN: OA12814.2 Add.

FID#: 14714

Notes:

- All fittings shall be registered in the name of the Manufacturer.
- Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
- The declaration shall be made by the person having full authority and responsibility for the quality of the end product.
- Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.



UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES

NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: Rosemount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92608	
PLANT LOCATIONS: Address above and Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY	
<p><input checked="" type="radio"/> A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers</p> <p><input type="radio"/> B Flanges: all flanges</p> <p><input type="radio"/> C Valves: all line valves</p> <p><input type="radio"/> D Expansion joints, flexible connections, and hose assemblies: all types</p> <p><input type="radio"/> E Strainers, filters, separators, and steam traps</p> <p><input type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters</p> <p><input type="radio"/> G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs</p> <p><input type="radio"/> H Pressure retaining components that do not fall into one of the above categories</p> <p><input type="radio"/> N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)</p>	<p>TITLE OF THE STANDARD OF CONSTRUCTION ASME B31.3 and proprietary standards, supported by proof pressure tests and calculations.</p>
<p>SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT</p> <p align="center">ROSEMOUNT[®] Analytical</p>	<p>TYPE OF CONSTRUCTION</p> <p>FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/> CAST <input type="checkbox"/> OTHER <input checked="" type="checkbox"/> DESCRIBE OTHER: Cast and machined</p>
<p>LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED:</p> <p>Drawings of main pressure retaining parts of kits 23160-00/01, 23311-00/01, 23760-00/01, and 23240-00 Report containing thread strength calculations for the main pressure retaining parts of the kits listed above Proof pressure reports RAI 01 rev. 2 and RAI 02 rev.2.</p>	

DECLARATION:

I, Dana Crowley (see note 3) employed by Rosemount Analytical, Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Del Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: [Signature]
 Declared before me at Irvine, County of Orange, California
 This 30th day of November AD 2013
 Commissioner of Oaths or Notary Public: (sign) [Signature]




This space for Regulatory Authority use

This registration must be revalidated after ten (10) years from the date of acceptance

CRN: OA12814.2 Add.

FID#: 14714

Notes:
 1. All fittings shall be registered in the name of the Manufacturer.
 2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
 3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product.
 4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.



REGISTERED

UNDER THE AUTHORITY OF THE BOILER AND PRESSURE VESSEL ACT.

C.R.N. OA12814.2T Add

SIGNED [Signature]

DATE Feb 17 2013

\$ 150.00

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS.

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES

NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: Rosemount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92606	
PLANT LOCATIONS: Address above and Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY	
<input checked="" type="radio"/> A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers <input type="radio"/> B Flanges: all flanges <input type="radio"/> C Valves: all line valves <input type="radio"/> D Expansion joints, flexible connections, and hose assemblies: all types <input type="radio"/> E Strainers, filters, separators, and steam traps <input type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters <input type="radio"/> G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs <input type="radio"/> H Pressure retaining components that do not fall into one of the above categories <input type="radio"/> N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> . (Meeting CNSC or ASME requirements)	TITLE OF THE STANDARD OF CONSTRUCTION ASME B31.3 and proprietary standards, supported by proof pressure tests and calculations.
SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT	TYPE OF CONSTRUCTION FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/> CAST <input type="checkbox"/> OTHER <input checked="" type="checkbox"/> DESIGNER OTHER Cast and machined
ROSEMOUNT Analytical	
LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED: Drawings of main pressure retaining parts of kits 23166-00/01, 23311-00/01, 23785-00/01, and 23240-00 Report containing thread strength calculations for the main pressure retaining parts of the kits listed above Proof pressure reports RAI 01 rev. 2 and RAI 02 rev. 2.	

DECLARATION:

I, Dana Crowley (see note 3) employed by Rosemount Analytical, Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Dal Norsa Vargas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: *Dana Crowley*
Declared before me at Irvine, County of Orange, California
This 30th day of November AD 2013
Commissioner of Oaths or Notary Public: (sign) *Catherine R. [Signature]*

U S
CATHERINE ANN STIP
Commission # 1822116
Notary Public - California
Orange County
My Comm. Exp. 02/21/2017

This space for Regulatory Authority use
This registration must be revalidated after ten (10) years from date of acceptance

Boilers and Pressure Vessels Act

CRN: 0A12814.2 Add.
FID#: 14714

Notes:
1. All fittings shall be registered in the name of the Manufacturer.
2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product.
4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.

REGISTERED
CRN 0A12814.2 N Add
Date FEB 01 2013
Signed *[Signature]*
Chief Inspector

Territorial Registration Fee \$1003

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK NOVA SCOTIA PRINCE EDWARD ISLAND NEWFOUNDLAND AND LABRADOR
 NUNAVUT YUKON NORTHWEST TERRITORIES

MANUFACTURERS NAME: Rosemount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92606	
PLANT LOCATIONS: Address above and Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY	
(A) Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers (B) Flanges: all flanges (C) Valves: all line valves (D) Expansion joints, flexible connections, and hose assemblies: all types (E) Strainers, filters, separators, and steam traps (F) Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters (G) Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs (H) Pressure retaining components that do not fall into one of the above categories (N) Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)	TITLE OF THE STANDARD OF CONSTRUCTION ASME B31.3 and proprietary standards, supported by proof pressure tests and calculations.
ROSEMOUNT[®] Analytical	TYPE OF CONSTRUCTION FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/> CAST <input type="checkbox"/> OTHER <input checked="" type="checkbox"/> DESCRIBE OTHER: Cast and machined
SHOW MANUFACTURER NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT	
LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED: Drawings of main pressure retaining parts of kits 23166-00/01, 23311-00/01, 23765-00/01, and 23240-00 Report containing thread strength calculations for the main pressure retaining parts of the kits listed above Proof pressure reports RAI 01 rev. 2 and RAI 02 rev.2.	

DECLARATION:

I, Dana Crowley (see note 2) employed by Rosemount Analytical, Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: *Dana Crowley*
 Declared before me at Irvine, County of Orange, California
 This 30th day of November AD 2012
 Commissioner of Oaths or Notary Public: (sign) *Catherine Ann Stip*



This space for Regulatory Authority use This registration must be revalidated after ten (10) years from the date of acceptance. CRN: <u>OA12814.2 Add.</u> FID#: <u>14714</u> Notes: 1. All fittings shall be registered in the name of the Manufacturer. 2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation. 3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product. 4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.	GOVERNMENT OF Newfoundland AND LABRADOR Labrador Department of Government Services Registered <u>OA12814.2 Add.</u> Date <u>2013/2/12</u> Registered by <u><i>[Signature]</i></u> UNDER THE AUTHORITY OF THE PUBLIC SERVICE ACT, 1988/2003 THE BOILER, PRESSURE VESSEL AND COMPRESSED GAS REGULATIONS
---	---