

GAS ANALYSERS

The Pressure Equipment Directive 97/23/EC


Daniel Europe Ltd hereby confirm that all gas analysers manufactured by them are exempt from the requirements of the above directive.

Daniel gas analysers conform to European Directive 94/9/EC (the ATEX Directive) and are less than category 1 as defined in Directive 97/23/EC (PED) Annex II table 6.

Therefore under paragraph 3.6 of article 1 of the directive 97/23/EC (printed below) the analysers are exempt from the requirements of that directive.

3. The following are excluded from the scope of this Directive :

Directive 97/23/EC Article 1 Para 3.6
<p>3.6. equipment classified as no higher than category I under article 9 of this Directive and covered by one of the following Directives: (See Guidelines: 1/12, 1/19, 1/26, 1/39, 1/48)</p> <ul style="list-style-type: none">- Council Directive 89/392/EEC of 14 June 1989 on the approximation of the laws of the Member States relating to machinery (5),- European Parliament and Council Directive 95/16/EC of 29 June 1995 on the approximation of the laws of the Member States relating to lifts (6),- Council Directive 73/23/EEC of 19 February 1973 on the harmonization of the laws of the Member States relating to electrical equipment designed for use within certain voltage limits (7),- Council Directive 93/42/EEC of 14 June 1993 concerning medical devices (8),- Council Directive 90/396/EC of 29 June 1990 on the approximation of laws of the Member States relating to appliances burning gaseous fuels (9),- Directive 94/9/EC of the European Parliament and the Council of 23 March 1994 on the approximation of laws of the Member States concerning equipment and protective systems intended for use in potentially explosive atmospheres (10);



(Neil Hendry)
GC Sustaining Engineer
Emerson Process Management