

1151 Boston-Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781 762 9375 www.fmglobal.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

1055a-b-c. Dual Input Conductivity/Resistivity Analyzer.

NI/1/2/ABCD/T6 Ta = 50°C; Type 4X

a= Span Limits blank, or C.

b= Voltage Rating 01 or 02.

c= Mounting 10, 11 or 12.

Equipment Ratings:

Nonincendive for use in Class I, Division 2, Groups A, B, C, and D; Temperature Class T6 Ta = 50°C; indoor and outdoor, Type 4X Hazardous (Classified) Locations.

Approved for:

Rosemount Analytical Incorporated
Irvine, CA 92606 USA

This certifies that the equipment described has been found to comply with the following Factory Mutual Research Approval Standards and other documents:


Class 3600	1998
Class 3611	1999
Class 3810	1989
ANSI/NEMA 250	1991

Original Approval Job Identification: 3004065 Approval Granted: January 23, 2001

Subsequent Revision Reports / Date Approval Amended

3010421	December 3, 2001
---------	------------------

Factory Mutual Research Corporation



Paris Stavrianidis
Vice President
Approvals Division

December 6, 2001
Date

An FM Global Affiliate