

Ex

Ex



British Approvals Service for Electrical
Equipment in Flammable Atmospheres

Certificate of Conformity Variation

THIS IS TO CERTIFY THAT CERTIFICATE BAS NO Ex 822245

Held by ROSEMOUNT ANALYTICAL INC
of Irvine, California 92714, USA

for the MODEL 1181 pH/ORP TRANSMITTER

is hereby extended to apply to apparatus designed and constructed in accordance with the specification set out in the Schedule of the said Certificate but having the variations specified in the following Schedule.

VARIATION FIFTEEN To permit the use of alternative terminal blocks which incorporate RFI filters. This change does not affect the original safety assessment.

DRAWINGS

Number	Issue	Date	Description
23563-000 (was 200252B)	A	5.21.93	S-Assy, HSG 2 Wire RMT
9120420/421	D	8.29.91	RFI Terminal Block

File No: EECS 0911/02/001

Certification Report No: 93(C)296 dated 24 September 1993



CERTIFICATE BAS NO Ex 822245/10

LP

I M CLEARE
DIRECTOR EECS
4 October 1993



Electrical Equipment Certification Service
Health and Safety Executive
Harpur Hill, Buxton, Derbyshire, SK17 9JN, United Kingdom
Tel: 0298 26211 Fax: 0298 79514 Telex: 668113 RLSD G



Ex

Ex