

British Approvals Service for Electrical
Equipment in Flammable Atmospheres



Certificate of Conformity - Variation

SUPPLEMENTARY CERTIFICATE BAS No. Ex 96D2494/7

This is to certify that Apparatus Certificate number:

Ex 96D2494

held by:

ROSEMOUNT ANALYTICAL INC

of

2400 Barranca Parkway, Irvine, California 92714-5018, USA

for the:


pH/ORP TRANSMITTER MODEL 3081

is hereby extended to apply to the apparatus designed and constructed in accordance with the specification set out in the Schedule of the said Certificate but having any variations specified in the Schedule attached to this certificate and the documents therein referred to.

This Supplementary Certificate shall be held with the original Certificate.

File No: EECS 0911/02/015

Sheet 1 of 2



This certificate is granted subject to the general conditions of the Electrical Equipment Certification Service. It does not necessarily indicate that the apparatus may be used in particular industries or circumstances. Representation of equipment as "Certified" is valid only when the number of the prime certificate to which this certificate is a supplement is given on the relevant EECS Manufacturing Licence or Verification Certificate.

**I M CLEARE
DIRECTOR
15 June 2001**



Electrical Equipment Certification Service
Health and Safety Executive
Harpur Hill, Buxton, Derbyshire, SK17 9JN, United Kingdom
Tel: +44 (0)1298 28000 Fax: +44 (0)1298 28244
internet: www.baseefa.com e-mail: baseefa.info.eecs@hsl.gov.uk



British Approvals Service for Electrical
Equipment in Flammable Atmospheres



Schedule

Supplementary Certificate BAS No. Ex 96D2494/7

VARIATION SEVEN

To permit:-

1. Component changes on the 81pH Analog circuit board.
2. Drawing information changes and corrections.

DRAWINGS

<u>Number</u>	<u>Issue</u>	<u>Date</u>	<u>Description</u>
*23817-00/-03 Sheets 1 & 2	E	1.27.01	4081 CPU Assembly
23832-00/-02	C	2.24.00	Fieldbus Output Assembly Mod 4081
23862-00/01	B	9.13.00	3081pH/4081pH Analog Assembly
2400292 Sheets 1 & 2	AF	5.22.01	81pH Analog Board Schematic
23637-00 Sheets 1 & 2	AF	5.24.01	81pH Analog Board Assembly

* This drawing was previously numbered 23817-00/-04