



Certificate of Conformity - Variation

This is to certify that Equipment Certificate number: **Ex 96D2494**
held by: **ROSEMOUNT ANALYTICAL INC**
of: **24000 Barranca Parkway, Irvine, California 92714-5018, USA**
for the: **pH/ORP TRANSMITTER MODEL 3081**

is hereby extended to apply to the apparatus designed and constructed in accordance with the specification set out in the Schedule of the said certificate but having any variations specified in the Schedule attached to this certificate and the documents therein referred to.

This Supplementary Certificate shall be held with the original Certificate

The Electrical Equipment Certification Service retains responsibility for its original documentation. Baseefa is responsible only for the additional work relating to this supplementary certificate and any other supplementary certificate that it has issued.

This certificate may only be reproduced in its entirety, without any change, Schedule included.

Baseefa Customer Reference No. **0911**

Project File No. **04/0693**

This certificate is granted subject to the general terms and conditions of Baseefa (2001) Ltd. It does not necessarily indicate that the equipment may be used in particular industries or circumstances.

R S SINCLAIR
DIRECTOR
On behalf of
Baseefa (2001) Ltd.

Baseefa
Rockhead Business Park, Staden Lane,
Buxton, Derbyshire SK17 9RZ
Telephone +44 (0) 1298 766600 Fax +44 (0) 1298 766601
e-mail info@baseefa.com web site www.baseefa.com
Baseefa is a trading name of Baseefa (2001) Ltd
Registered in England No. 4305578 at the above address



Schedule

Description of the variation

Variation 9.1

To permit the Analogue, CPU and Fieldbus PCBs to be replaced by the equivalent versions from the Model 5081-A Transmitter (Certificate No. BAS02ATEX1284), Model 5081-C Transmitter (Certificate No. Baseefa03ATEX0099) and Model 5081-G Transmitter (Certificate No. Baseefa04ATEX0052X).

Input/output parameters are unchanged.

Report No.

None

Drawings

Number	Sheet	Issue	Date	Description
1700425*	1 to 3	D	9-12-02	Certified Product Schematic, 5081- A/P/C/T- CPU
1700426*	1 & 2	A	4-23-02	Certified Product Assy, PCB, 5081-A/P/C/T CPU
1700427 ^{##}	1	A	4-23-02	Certified Product Assy, PCB, 5081-pH HT ANALOG
1700428 ^{##}	1 to 3	D	05-20-02	Certified Product Schematic, PCB, 5081-A/P HT/FF Analog
1700478 [#]	1	B	10-17-03	Certified Product Assy, 5081-G Analog Board
1700479 [#]	1 to 2	B	02-03-04	Certified Product, Schematic, 5081-G Analog Board
1700549*	1	A	5-6-04	Certified Product Assy, 5081-A/P/C/T-FF Fieldbus CPU
1700550*	1 to 3	A	5-6-04	Certified Product SCH, 5081- A/P/C/T-FF-Fieldbus CPU
33718-00 ^{##}	1	E	5-3-02	PCB, 5081-A/P Analog Board
33773-00*	1	B	5-3-02	PCB, 5081-A/P/C/T CPU
33885-00 [#]	1	A	07.08.03	PCB, 5081-G Analog
33890-00*	1	A	5-6-04	PCB, CPU, 5081-A/P/C/T-FF/FI
70025-00**	1	B	5-6-04	PCB, Fieldbus Square Card

* - These drawings are held with Baseefa03ATEX0099

** - This drawing is common to and held with Ex97D2268X/8

- These drawings are held with Baseefa04ATEX0052X

- These drawings are held with BAS02ATEX1284

Special Conditions for Safe Use

None.