



Certificate of Compliance

Certificate: 2447828 (LR 39156)

Master Contract: 226885

Project: 2548848

Date Issued: February 15, 2013

Issued to: Rosemount Analytical, Inc.
10241 W. Little York, Suite 200
Houston, TX 77040
USA
Attention: Vicente Ramirez

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Marin Banu

Issued by: Marin Banu, P. Eng.

PRODUCTS

CLASS 2258 02 - PROCESS CONTROL EQUIPMENT - For Hazardous Locations

CLASS 2258 82 - PROCESS CONTROL EQUIPMENT - For Hazardous Locations -
Certified to US Standards

Class I, Division 1, Groups B, C, and D; Encl. Type 4X

Class I, Zone 1, Ex/AEx d IIB + H2, T6; IP65.

- Gas Chromatograph System, Model 370XA, rated 21-30 V d.c., max. 55 W, Ambient temperature; -20°C to 60°C, Temp Code T6.

APPLICABLE REQUIREMENTS

CSA Std. C22.2 No. 0-10

- General Requirements Canadian Electrical Code, Part II

CAN/CSA Std. C22.2 No. 94.1-07

- Enclosures for Electrical Equipment, Non-Environmental Considerations



Certificate: 2447828 (LR 39156)

Master Contract: 226885

Project: 2548848

Date Issued: February 15, 2013

-
- | | |
|---|--|
| CAN/CSA Std. C22.2 No. 94.2-07 | - Enclosures for Electrical Equipment, Environmental Considerations |
| CAN/CSA Std. C22.2 No. 61010-1-04 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |
| CSA Std. C22.2 No. 30-M1986 (R 2003) | - Explosion-Proof Enclosures for Use in Class I Hazardous Locations |
| CAN/CSA Std. C22.2 No. 60079-0:07 | - Electrical apparatus for explosive gas atmospheres – Part 0: General requirements |
| CAN/CSA Std. C22.2 No. 60079-1:07 | - Electrical apparatus for explosive gas atmospheres – Part 1: Flameproof enclosures "d" |
| CAN/CSA Std. C22.2 No. 60529:05 | - Degrees of protection provided by enclosures (IP Code) |
| ANSI/UL Std. 61010-1-2004 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |
| ANSI/UL Std. 50 | - Enclosures for Electrical Equipment, Non-Environmental Considerations |
| ANSI/UL Std. 50E | - Enclosures for Electrical Equipment, Environmental Considerations |
| ANSI/UL Std. 1203 | - Explosion-Proof and Dust-Ignition-Proof Electrical Equipment for Use in Hazardous (Classified) Locations |
| ANSI/UL Std. 60079-0 (5th Edition 2009) | - Explosive Atmospheres – Part 0: Equipment - General Requirements |
| ANSI/UL 60079-1 Std. (6th Edition 2009) | - Explosive Atmospheres – Part 1: Equipment Protection by Flameproof Enclosures “d” |
| ANSI/IEC Std. 60529:2004 | - Degrees of protection provided by enclosures (IP Code) |



Supplement to Certificate of Compliance

Certificate: 2447828

Master Contract: 226885

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
2548848	Feb 15, 2013	Update to report 2447828 to add new window material (polycarbonate) and potting epoxy.
2447828	Jun 15, 2012	CSAc-us Certification of 370XA Process Gas Chromatograph for Class I, Div 1, Grps B, C, D & Class I, Zone 1, Ex/AEx d IIB + H2; T6. Type 4X, IP65.