

British Approvals Service for Electrical
Equipment in Flammable Atmospheres



Certificate of Conformity - Variation

SUPPLEMENTARY CERTIFICATE BAS No. Ex 97D2305X/1

This is to certify that Apparatus Certificate number:

Ex 97D2305X

held by:

ROSEMOUNT ANALYTICAL INC

of:

2400 Barranca Parkway, Irvine, California 92714-5018, USA

for the:

pH/ORP SENSOR MODEL 396P-10/12-50


is hereby extended to apply to the apparatus designed and constructed in accordance with the specification set out in the Schedule of the said Certificate but having any variations specified in the Schedule attached to this certificate and the documents therein referred to.

This Supplementary Certificate shall be held with the original Certificate.

File No: EECS 0911/02/018

Sheet 1 of 2

This certificate is granted subject to the general conditions of the Electrical Equipment Certification Service. It does not necessarily indicate that the apparatus may be used in particular industries or circumstances. Representation of equipment as "Certified" is valid only when the number of the prime certificate to which this certificate is a supplement is given on the relevant EECS Manufacturing Licence or Verification Certificate.



**I M CLEARE
DIRECTOR**

16 November 1999



Electrical Equipment Certification Service
Health and Safety Executive
Harpur Hill, Buxton, Derbyshire, SK17 9JN, United Kingdom
Tel: 01298 28000 Fax: 01298 28244



Registration Number
020
The use of the Accreditation
Mark indicates accreditation in
respect of those activities
covered by the accreditation
certificate number 020.

British Approvals Service for Electrical
Equipment in Flammable Atmospheres



Schedule

Supplementary Certificate BAS No. Ex 97D2305X/1

VARIATION ONE

To permit the following:-

1. The use of an alternative printed circuit board fitted with surface mount components.
2. Modifications to a number of drawings to change production related information.

The above changes are considered not to affect the intrinsic safety assessment.

The Special Conditions for Safe Use specified on the original certificate remain applicable.

DRAWINGS

<u>Number</u>	<u>Sheet</u>	<u>Issue</u>	<u>Date</u>	<u>Description</u>
396P	1 to 4	P	9.15.99	General assembly 369P
23670-00/62	1 to 4	R	1.11.99	Sub assembly
2400205		E	11.3.99	Circuit - pre-amplifier Model 399
22941-00		H	10.13.99	PCB assembly Model 399
32790-00		F	9.28.99	PCB details Model 399