



# EU Declaration of Conformity



(No. 1700906)

5081-P/A...

This declaration is issued under the sole responsibility of the manufacturer:  
Rosemount Inc., 8200 Market Blvd., Chanhassen, MN 55317 USA

The product,

## Rosemount Amperometric/pH Transmitter model 5081-AA-BB-CC-DD

Where

AA is:

P pH/ORP Measurement  
A Amperometric (oxygen, ozone, chlorine)

CC is:

20 Infrared remote controller included  
21 Infrared remote controller not included

BB is:

HT Analog/HART communications  
FF Fieldbus communications  
FI FISCO communications

DD is:

60 No approval  
67 FM, Intrinsicly Safe, NonIncendive, Explosion-proof, Dust Ignition proof  
69 CSA, Intrinsicly Safe, NonIncendive, Explosion -proof  
73 ATEX / IECEx, Intrinsicly Safe

to which this declaration relates, is in conformity with relevant Union harmonization legislation:

(2014/30/EU) EMC Directive

(2014/34/EU) ATEX Directive (The ATEX Directive is only valid if option 73 is selected)

Provisions of the directive fulfilled by the equipment:

Equipment Group II, Category 1 G Ex ia IIC T4 Ga (-20°C ≤ Ta ≤ +65°C)

Intrinsicly Safe Certificate: BAS02ATEX1284X

Special Condition for safe use:

The equipment enclosure may contain light metals. The apparatus must be installed in such a manner as to minimize the risk of impact or friction with other metal surfaces.

ATEX Notified Body for EC Type Examination Certificate & Quality Assurance:

SGS Baseefa[Notified Body Number:1180], Rockhead Business Park, Staden Lane, Buxton SK17 9RZ UNITED KINGDOM

Assumption of conformity is based on the application of the harmonized standards:

EN 61326-1:2006 Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements

EN 60079-0:2012+A11:2013 Explosive atmospheres. Equipment. General requirements

EN 60079-11:2012 Explosive atmospheres. Equipment protection by intrinsic safety "i"

(Signature)

Kim Freeman  
(Name printed)

Director of Global Quality  
(Function name)

March 23, 2017  
(Date of issue)