



Member of the FM Global Group

FM Approvals  
1151 Boston Providence Turnpike  
P.O. Box 9102 Norwood, MA 02062 USA  
T: 781 762 4300 F: 781-762-9375 www.fmapprovals.com

# CERTIFICATE OF COMPLIANCE

## HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

***IRC (Infrared Remote Control)\*. Handheld IR Transmitter***

IS / I / 1 / ABCD / T4 Ta = -20 °C to 50 °C

I / 0 / AEx ia / IIC / T4 Ta = -20 °C to 50 °C

*\*Part Number: 24479-00*

*Special Condition for use:*

- 1. Use only ENERGIZER E92/EN92 or DURACELL MN2400/PC2499 batteries*

### Equipment Ratings:

Intrinsically Safe for Class I, Division 1, Groups A, B, C and D hazardous (classified) locations,  
Intrinsically Safe for Class I, Zone 0, Group IIC hazardous (classified) locations

### FM Approved for:

Emerson Process Mngmt-Rosemount Analytic  
Irvine, CA United States



Member of the FM Global Group

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

Class 3600	2011
Class 3610	2010
Class 3810	2005
ANSI/ISA 61010-1	2012
ANSI/ISA 60079-0	2013
ANSI/ISA 60079-11	2012

Original Project ID: 0003051016

Approval Granted: August 14, 2014

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC

James E. Marquedant  
Manager, Electrical Systems

14 August 2014

Date