



CSA INTERNATIONAL

Certificate of Compliance

Certificate: 1001031 (LR 8692)

Master Contract: 156421

Project: 1001031 (Edition 1)

Date Issued: September 26, 2000

Issued to: Rosemount Analytical Inc.
4125 E. La Palma Avenue
Anaheim, CA 92807-1802
USA

The products listed below are eligible to bear the CSA Mark shown, with adjacent indicator "C" and "US".



Issued by: Dan Livermore, C.E.T.

Signature: _____

CLASS

2252 01 - PROCESS CONTROL EQUIPMENT

2252 83 - PROCESS CONTROL EQUIPMENT - CERTIFIED TO U.S. STANDARDS

PRODUCTS

Model MCFID, Multi Component Flame Ionization Detector, Model Part No. 194009, rated 24V, 120W max. and Model FID2 Multi Component Flame Ionization Detector, Model Part No. 194109, rated 24V, 120W max.

Notes:

1. The subject equipment is intended for use in ordinary non-hazardous locations only. Inherent in the design of the equipment are methods of protection intended to provide safety in the event of a self-generated hazard such as the internal release of flammable gases.
2. The equipment is certified as a component, intended for rack or panel mounting. An appropriate enclosure is to be provided in the end use application according to the manufacturer's instructions. The suitability of the final combination to be determined by the authority having jurisdiction.

The "C" and "US" indicator adjacent to the CSA Mark signifies that the product has been evaluated to the applicable ANSI/UL and CSA Standards, for use in the U.S. and Canada. This includes products eligible to bear the NRTL indicator. NRTL, i.e. Nationally Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.