



**Bureau Veritas E&E  
Product Services GmbH**  
Businesspark A96  
86842 Türkheim  
Deutschland  
+ 49 (0) 8245 96810-0  
info-tur@de.bureauveritas.com

## Prüfbescheinigung Certificate

**Antragsteller** Emerson Process Management GmbH&Co. OHG  
Applicant Industriestrasse 1  
63594 Hasselroth  
Germany

**Produkttyp** Laboratory Equipment  
Product type

**Modell** **X-STREAM** Field Housing Gas Analyzer (X2F)  
Model **X-STREAM** General Purpose Gas Analyzer (X2GP)  
**X-STREAM** Gas Analyzer Core (X2CA)  
**X-STREAM** General Purpose Compact Gas Analyzer (X2GC)  
**X-STREAM** Compact Gas Analyzer Core (X2CC)

**Leistung** 100 – 240Vac; 50/60Hz; 3 – 1,5A (X2F, X2GP, X2CA)  
Ratings 24Vdc; 2,5A (X2GC, X2CC)

Ein repräsentatives Testmuster des o.g. Modells bestand die Prüfung nach  
A representative Test sample of above stated model passed the tests according  
to

**Norm** IEC 61010–1:2001 (2<sup>nd</sup> Edition); EN 61010–1:2001 (2<sup>nd</sup> Edition);  
Standard DIN EN 61010–1:2002; UL 61010–1 (2<sup>nd</sup> Edition);  
CAN/CSA C22.2 No. 61010–1:2004

**Berichtsnr.** 05KFS106, Rev.5  
Report No.

**Zertifikat Nr.** 07-195  
Certificate No.

**Datum** 13.12.2007  
Issued

**Zertifizierungsstelle**  
Certification department

  
A. Hänchen