



# 防爆合格证

证号: GYJ12.1025X

由 艾默生过程控制有限公司

制造的产品:

(地址: 上海浦东金桥出口加工区新金桥路1277号)

名称 分析仪

型号规格 X-STREAM a

防爆标志 Ex d II B+H<sub>2</sub> T4 Gb

产品标准

图样编号 4.271-7112/1

经图样及技术文件的审查和样品检验, 确认上述产品符合 GB 3836.1-2010、GB 3836.2-2010 标准, 特颁发此证。

本证书有效期: 2012年3月5日至2017年3月4日

- 备注
1. 安全使用注意事项见本证书附件。
  2. 证书编号后缀“X”表明产品具有安全使用特殊条件: 涉及隔爆接合面的维修须联系制造商。
  3. 型号规格说明: 型号规格中a代表不同电气配置, 可为FD、1FD、2FD、100FD或EFD。

站长

国家级仪器仪表防爆安全监督检验站

颁发日期二〇一二年三月五日

本证书仅对与认可文件和样品一致的产品有效。

地址: 上海市漕宝路103号  
邮编: 200233

网址: [www.nepsi.org.cn](http://www.nepsi.org.cn)  
Email: [info@nepsi.org.cn](mailto:info@nepsi.org.cn)

电话: +86 21 64368180  
传真: +86 21 64844580



# EXPLOSION PROTECTION CERTIFICATE OF CONFORMITY

Cert NO.GYJ12.1025X

This is to certify that the product

Analyzer

manufactured by Emerson Process Management Co., Ltd.

(Address:No.1277, Xinqinqiao Road, Pudong, Shanghai)

which model is X-STREAM a

Ex marking Ex d II B+H<sub>2</sub> T4 Gb

product standard /

drawing number 4.271-7112/1

has been inspected and certified by NEPSI, and that it conforms to GB 3836.1-2010,GB 3836.2-2010

This Approval shall remain in force until 2017.03.04

Remarks 1.Conditions for safe use are specified in the attachment to this certificate.  
2.Symbol "X" placed after the certification number denotes specific conditions of use: contact the original

manufacturer when the repair work relates to the flamepath.

3.Model designation: the symbol a in the type code represents different electrical configurations which could

Director



National Supervision and Inspection Centre for  
Explosion Protection and Safety of Instrumentation

Issued Date 2012.03.05

This Certificate is valid for products compatible with the documents and samples approved by NEPSI.

103 Cao Bao Road  
Shanghai 200233, China

<http://www.nepsi.org.cn>  
Email: info@nepsi.org.cn

Tel: +86 21 64368180  
Fax: +86 21 64844580