



## **Prüfbescheinigung** *Certificate*

**Antragsteller:** **Emerson Process Management GmbH & Co. OHG**  
*Applicant:* Industriestraße 1, 63594 Hasselroth – Germany

**Produkt Typ:** **Gas Analyzer**  
*Product type:*

**Modell:** **X - stream GP or GPS or F**  
*Model:* F...field housing  
GP...table top or rack mounted version with terminals  
GPS...table top or rack mounted version with sockets

Ein repräsentatives Testmuster des o.g Models bestand die Prüfung nach  
*A representative Test Sample of above stated model passed the tests according to*

**Norm:** IEC 61010 – 1 : 2001 (2<sup>nd</sup> Edition); EN 61010 – 1 : 2001 (2<sup>nd</sup> Edition);  
*Standard:* DIN EN 61010 – 1 : 2002; UL 61010 – 1 : 2<sup>nd</sup> edition;  
CAN/CSA C22.2 No. 61010 – 1 : 2004

**Technische Daten:** 100 – 240Vac; 50/60Hz; 2 – 1A  
*Ratings:*

**Bericht Nr:** **05KFS106**  
*Report No.:*

**Zertifikat Nr.:** **06 - 050**  
*Certificate No.:*

**Datum:** **05. April 2006**  
*Issued:*

**Zertifizierstelle: Horst Haug**  
*Certificate Placement: Horst Haug*