



- ① THE ANCHOR WEIGHT MUST BE KEPT CLEAR INSIDE THE PIPE.
- ② FOR CLEAN PRODUCTS ONLY. MIN PIPE SIZE FOR HIGH VISCOSITY PRODUCTS IS 3" SCH 40-80. THE WLS IS NOT RECOMMENDED TO BE INSTALLED IN STILL PIPE FOR WAXY CRUDE OILS.

| | | | | | | | |
|------------------|------|------------------|-------|------------------------|----------------------------|--------------|-------|
| ISSUED BY | WEEK | PRODUCT CODE | FILE | ACAD | INSTALLATION DRAWING TITLE | ISSUE | SHEET |
| GSP-AJ | 0243 | REX | ACAD | MECH. INSTALLATION DWG | 2 | 2 | 1/1 |
| APPROVED BY | WEEK | ORIGINAL DWG NO. | SCALE | WLS IN STILL PIPE | | | |
| GSP | 0243 | 9261 090-154 | - | | | | |
| ROSEMOUNT | | | | 1 ST ANGLE | DOC TYPE | DWG NO. | |
| Tank Gauging | | | | 1/4" | 2 | 9261 090-154 | |

THE COMPANY ASSUMES NO RESPONSIBILITY FOR THE DESIGN OR CONSTRUCTION OF THE TANK. THE CONSULTANT HAS NOT BE USED WITHOUT OUR AUTHORIZATION OR ENDORSEMENT TO THE KNOWLEDGE OF A THIRD PARTY. COMMUNICATION WILL BE PROSECUTED. ROSEMOUNT TANK GAUGING AB, SWEDEN.