

# Parts Certificate

Number: **TC8218** revision 11  
Project number: 1901358  
Page: 1 of 1

Issued by NMI Certin B.V.

In accordance with – WELMEC guide 8.8 "General and Administrative Aspects of the Voluntary System of Modular Evaluation of Measuring instruments under the MID".  
– OIML R117-1 Edition 2007 (E) "Dynamic measuring systems for liquids other than water".

Producer Emerson Process Management  
Remote Automation Solutions  
Unit 2  
Park Farm Courtyard  
Easthorpe, Malton  
North Yorkshire YO17 6QX  
United Kingdom

Part An **electronic calculating and indicating device** for use as part of a liquid measurement installation.

Producer's mark or name : Emerson  
Type designation : FloBoss S600 Flow computer or S600+ Flow computer  
Software versions : see paragraph 1.2.1 of the description  
Accuracy class : 0,3 and higher  
Environmental classes : M1 / E2  
Temperature range ambient : -10 °C ... +55 °C  
Humidity environment class : Non condensing

Further properties and test results are described in:

- Description number TC8218 revision 11;
- Documentation folder number TC8218-6.

Remark This revision replaces the previous revision.  
The documentation folder replaces the previous documentation folder.

Issuing Authority **NMI Certin B.V.**  
7 August 2017



C. Oosterman  
Head Certification Board